



Brian R. Leahy
Director

MEMORANDUM

Edmund G. Brown Jr.
Governor

TO: Joy Dias
Senior Environmental Scientist
Environmental Monitoring Branch

FROM: Vaneet Aggarwal
Environmental Scientist
916-445-5393

Original Signed By

DATE: July 26, 2017

SUBJECT: THE QUALIFICATION OF METHOD EM-5.5 AS UNEQUIVOCAL
ACCORDING TO CRITERIA IN THE PESTICIDE CONTAMINATION
PREVENTION ACT

BACKGROUND

The Pesticide Contamination Prevention Act (Food and Agricultural Code [FAC] sections 13141 et seq.) was passed in 1985 to prevent further pesticide pollution of ground water which may be used for drinking water supplies. FAC section 13149 specifies the conditions under which a pesticide is considered “found” in ground water or soil, and thus subject to formal review as specified. FAC subsection 13149(d) allows a finding of a pesticide in ground water or soil to be based on a single analytical method conducted by a single analytical laboratory, only if the analytical method provides unequivocal identification of a chemical. Criteria to identify methods providing unequivocal identification of a chemical are included in a DPR memo entitled “Evaluating analytical methods for compliance with the Pesticide Contamination Prevention Act Requirements” (Aggarwal, 2012).

PURPOSE

Determine if the analytical method (EM-5.5) for Bentazon in well water used by the California Department of Food and Agriculture (CDFA) meets the definition of an unequivocal detection method.

DISCUSSION AND RECOMMENDATION

CDFA Center for Analytical Chemistry method (EM-5.5) uses a gas chromatograph equipped with a mass selective detector (GC/MSD) for the detection of Bentazon in well water. Prior to injection of sample into the GC/MSD, ground water sample is acidified to pH less than 2 and then extracted with methyl tertiary butyl ether.

In CDFA method EM-5.5 for Bentazon, the MSD is run in selected ion monitoring mode looking only for selected molecular ion mass and specific daughter ion masses at specific retention times.

In CDFA method EM-5.5, the following criteria are used to confirm the presence of Bentazon in well water:

1. Each set of samples will have a matrix blank and a spiked matrix sample.
2. The retention time should be within ± 2 percent of that of the standards.
3. The recoveries of the matrix spikes shall be within the control limits.
4. The sample must be diluted if results fall outside the linear range of the standard curve.
5. Bracketing standard curves should have a percentage change less than 20%.
6. The R^2 of each calibration curve should be equal to or greater than 0.995.
7. The matrix blank should be free of target compounds above the reporting level.

Analysis of Bentazon by method EM-5.5 is highly specific and qualifies for unequivocal detection designation. Therefore, analysis by a second laboratory or a second method is not necessary for well water samples analyzed for Bentazon by this method.

APPROVED *Original Signed by* Date *7-26-17*
Joy Dias
Senior Environmental Scientist

APPROVED *Original Signed by* Date *7-26-17*
Sheryl Gill
Environmental Program Manager I

APPROVED *Original Signed by* Date *7-26-17*
Pamela Wofford,
Environmental Program Manager II

Joy Dias
July 26, 2017
Page 3

REFERENCES

Aggarwal, V. 2012. Memorandum to Lisa Ross, Ph.D. Evaluating analytical methods for compliance with the Pesticide Contamination Prevention Act requirements.
http://www.cdpr.ca.gov/docs/emon/pubs/ehapreps/analysis_memos/2391_ross.pdf (accessed 20 March 2016).